


CAS NO.: 104264-80-4, CDCL3



Parameter	Value
1 Title	
2 Spectrometer magnet	
3 Solvent	CDCl3
4 Temperature	295.4
5 Pulse Sequence	s1pul30
6 Number of Scans	64
7 Receiver Gain	64
8 Relaxation Delay	1.0000
9 Pulse Width	9.8000
10 Acquisition Time	3.9999
11 Acquisition Date	2021-06-16T21:15:02
12 Modification Date	2021-06-16T21:33:08
13 Spectrometer Frequency	400.00
14 Spectral Width	8108.0
15 Lowest Frequency	-1654.0
16 Nucleus	1H
17 Acquired Size	32431
18 Spectral Size	65536