



Parameter	Value
1 Title	BIZ2020-N12-SCH1
2 Spectrometer	magnet
3 Solvent	DMSO
4 Temperature	296.8
5 Pulse Sequence	spul30
6 Number of Scans	64
7 Receiver Gain	64
8 Relaxation Delay	1.0000
9 Pulse Width	9.8000
10 Acquisition Time	3.9999
11 Acquisition Date	2020-11-12T19:26:22
12 Modification Date	2020-11-12T19:43:58
13 Spectrometer Frequency	400.00
14 Spectral Width	8108.0
15 Lowest Frequency	-1654.0
16 Nucleus	1H
17 Acquired Size	32431
18 Spectral Size	65536