



Parameter	Value
1 Title	BIZ2020-N19-SCH2
2 Spectrometer	magnet
3 Solvent	DMSO
4 Temperature	297.9
5 Pulse Sequence	slpul30
6 Number of Scans	64
7 Receiver Gain	64
8 Relaxation Delay	1.0000
9 Pulse Width	9.8000
10 Acquisition Time	3.9999
11 Acquisition Date	2020-11-19T20:56:16
12 Modification Date	2020-11-19T21:14:00
13 Spectrometer Frequency	400.00
14 Spectral Width	8108.0
15 Lowest Frequency	-1654.0
16 Nucleus	1H
17 Acquired Size	32431
18 Spectral Size	65536