

<b>Analysis Item:</b>	<b>Optical Rotation (OR)</b>
<b>Instrument Code:</b>	<b>OR-02</b>
<b>Instrument:</b>	<b>Shenguang WZZ-2B Auto Polarimeter</b>
<b>Sample ID:</b>	<b>LJC202102230037</b>
<b>CAS:</b>	<b>154445-77-9</b>
<b>Test Date:</b>	<b>2021.02.23</b>
<b>Operator:</b>	<b>SYW</b>

<b>Specification:</b>	<b>C=1 (DMF) -5.5</b>			
<b>Method:</b>	<b>Temp.[°C]</b>	<b>Solvent</b>	<b>Conc.[g/ml]</b>	<b>WL[nm]</b>
	<b>Room Temperature</b>	<b>DMF</b>	<b>0.010155</b>	<b>589</b>
<b>Result:</b>	<b>Arc[°]</b>		<b>OROT[°]</b>	
	<b>-0.060</b>		<b>-5.9</b>	
<b>Checker:</b>	<b>GYG</b>			