

Analysis Item:	Optical Rotation (OR)
Instrument Code:	OR-03
Instrument:	Shenguang WZZ-2B Auto Polarimeter
Sample ID:	LJC20210531013
CAS ID:	587-33-7
Test Date:	2021.05.31
Operator:	CC

Specification:	Ref: -7.9°			
	(C=1 1N HC)			
Method:	Temp.[°C]	Solvent	Conc.[g/ml]	WL[nm]
	Indoor Temperature	HCL	0.010144	589
Result:	Arc[°]		OROT[°]	
	-0.088		-8.675	
Checker:	LWW			