

<b>Analysis Item:</b>	Optical Rotation (OR)
<b>Instrument Code:</b>	OR-04
<b>Instrument:</b>	MCP-5100
<b>Sample ID:</b>	LJC202107070180
<b>CAS:</b>	587-33-7
<b>Test Date:</b>	20210707
<b>Operator:</b>	Chongwei

<b>Specification:</b>	(C=1g/100mL,1N HCL,589nm) -7.9°			
<b>Method:</b>	<b>Temp.[°C]</b>	<b>Solvent</b>	<b>Conc.[g/ml]</b>	<b>WL[nm]</b>
	20	1N HCL	1.0059	589
<b>Result:</b>	<b>Arc[°]</b>		<b>OROT[°]</b>	
	-0.096		-9.5	
<b>Checker:</b>	Wangzhuhai			