

Analysis Item:	Optical Rotation (OR)
Instrument Code:	OR-03
Instrument:	Shenguang WZZ-2B Auto Polarimeter
Sample ID:	LJC202106090201
CAS ID:	59279-60-6
Test Date:	2021.06.10
Operator:	HYX

Specification:	Ref: 14.8°			
	(C=1 CHCL3)			
Method:	Temp.[°C]	Solvent	Conc.[g/ml]	WL[nm]
	Indoor Temperature	CHCL3	0.009997	589
Result:	Arc[°]		OROT[°]	
	+0.160		+16.00	
Checker:	LW			