

| | |
|-------------------------|--|
| Analysis Item: | Optical Rotation (OR) |
| Instrument Code: | OR-03 |
| Instrument: | Shenguang WZZ-2B Auto polarimeter |
| Sample ID: | LJC2020-0509-1785 |
| Compound ID: | CS-W008464 |
| Method ID: | CS-W008464-H01OR(OR) |
| Test Date: | 20200509 |
| Operator: | ZLL |

| | | | | |
|-----------------------|-------------------------|----------------|--------------------|---------------|
| Specification: | Ref: -37° | | | |
| | (1g/100ml CHCL3) | | | |
| Method: | Temp.[°C] | Solvent | Conc.[g/ml] | WL[nm] |
| | 18 | CHCL3 | 0.010129 | 589 |
| Result: | Arc[°] | | OROT[°] | |
| | -0.378 | | -37.1 | |
| Checker: | WS | | | |