

<b>Analysis Item:</b>	<b>Optical Rotation (OR)</b>
<b>Instrument Code:</b>	<b>OR-03</b>
<b>Instrument:</b>	<b>Shenguang WZZ-2B Auto Polarimeter</b>
<b>Sample ID:</b>	<b>LJC2020-1009-0034</b>
<b>Compound ID:</b>	<b>CS-W016567</b>
<b>Method ID:</b>	<b>CS-W016567 -H01OR(OR)</b>
<b>Test Date:</b>	<b>20201009</b>
<b>Operator:</b>	<b>ZLL</b>

<b>Specification:</b>	<b>Ref: -14.2°</b>			
	<b>(1g/100ml H2O )</b>			
<b>Method:</b>	<b>Temp.[°C]</b>	<b>Solvent</b>	<b>Conc.[g/ml]</b>	<b>WL[nm]</b>
	<b>23</b>	<b>H2O</b>	<b>0.010069</b>	<b>589</b>
<b>Result:</b>	<b>Arc[°]</b>		<b>OROT[°]</b>	
	<b>-0.158 -15.7</b>			
<b>Checker:</b>	<b>XJ</b>			