

<b>Analysis Item:</b>	Optical Rotation (OR)
<b>Instrument Code:</b>	OR-03
<b>Instrument:</b>	Shenguang WZZ-2B Auto Polarimeter
<b>Sample ID:</b>	LJ2019-0617-0093
<b>Compound ID:</b>	CS-0643
<b>Batch No:</b>	Lc0617059
<b>Method ID:</b>	CS-0643-H01OR(OR)
<b>Test Date:</b>	20190618
<b>Operator:</b>	FJL
<b>Tracking:</b>	Lc0617059

<b>Specification:</b>	Ref: 71.1			
	(1g/100ml MEOH)			
<b>Method:</b>	Temp.[°C]	Solvent	Conc.[g/ml]	WL[nm]
	Indoor Temperature	MEOH	0.010012	589
<b>Result:</b>	Arc[°]		OROT[°]	
	0.726		72.5	
<b>Checker:</b>	GHJ			