

<b>Analysis Item:</b>	Optical Rotation (OR)
<b>Instrument Code:</b>	OR-03
<b>Instrument:</b>	Shenguang WZZ-2B Auto polarimeter
<b>Sample ID:</b>	LJC2020-0509-1785
<b>Compound ID:</b>	CS-W008464
<b>Method ID:</b>	CS-W008464-H01OR(OR)
<b>Test Date:</b>	20200509
<b>Operator:</b>	ZLL

<b>Specification:</b>	Ref: -37°			
	(1g/100ml CHCL3)			
<b>Method:</b>	Temp.[°C]	Solvent	Conc.[g/ml]	WL[nm]
	18	CHCL3	0.010129	589
<b>Result:</b>	Arc[°]		OROT[°]	
	-0.378		-37.1	
<b>Checker:</b>	WS			