

Analysis Item:	Optical Rotation (OR)
Instrument Code:	OR-03
Instrument:	Shenguang WZZ-2B Auto Polarimeter
Sample ID:	LJC2020-1030-0113
CAS ID:	4144-02-9
Test Date:	20201102
Operator:	ZLL

Specification:	Ref: -37 °			
	(1g/100ml H2O)			
Method:	Temp.[°C]	Solvent	Conc.[g/ml]	WL[nm]
	23°C	H2O	0.01003	589
Result:	Arc[°]		OROT[°]	
	-0.362		-36.1	
Checker:	XJ			