

File ..018\20180801\20180801\CPK2018-801-23585_3.D Tgt Mass (CHM):
Injection Date : 1 Aug 18 12:58 pm +0800 Seq. Line : 53
Sample Name : CPK2018-801-23585 Location : P1-D-10
Acq. Operator : SYSTEM Inj : 1
Spec. Reported : MS Integration Inj Volume : 5 ul
Acq. Method : D:\LCMS03\Data\2018\20180801\20180801\1-POS-10MIN.M
Analysis Method : D:\LCMS03\Data\2018\20180801\20180801\1-POS-10MIN.M
Catalog No : CS-4272 Batch#23585
Method Info : Mobile Phase: A: water(0.01%TFA) B:ACN(0.01%TFA)
Gradient: 5% to 95%B within 5 min
Flow Rate :1.8ml/min
Column :SunFire C18, 4.6*50mm,3.5um A-RP-186
Oven Temperature : 45C

ELS1 A, ELSD Signal



Integration Results for ELS1 A, ELSD Signal

RetTim	Width	Area	Height	Area%
0.39	0.05	3.20	0.89	0.41
2.96	0.05	0.42	0.14	0.05
3.02	0.07	0.29	0.02	0.04
3.19	0.06	784.03	220.13	99.31
3.34	0.04	0.81	0.33	0.10
3.47	0.04	0.71	0.21	0.09

*** End of Report ***