


CAS NO. : 3785-24-8

---

### HPLC

Chromatographic column: C18 column, 4.6 × 250mm × five µm  
Mobile phase: A: acetonitrile, B: 0.1% pa water, 0-20min, a: 40% - 80%  
Wavelength:  $\lambda = 264\text{nm}$   
Flow rate: 1ml / min  
Temperature: 30 °C  
Sample dissolution: methanol dissolution.



	RetTime	Area	%Area
1	16.674	14785652	99.50
2	20.148	71516	0.48
3	23.031	2890	0.02