

## 1 : HPLC

Catalog No: CS-6033 Batch#37349

Instrument: Agilent HPLC;

Chromatographic Column: Ultimate® XB-C18, 4.6\*250 mm, 5µm;


Moving Phase: MeOH-Water 10: 90;

Flow Velocity: 1.0 ml/min;

Column Temperature: 35 °C;

Detector: UV-210 nm

Configure the Concentration: 0.5mg/ml(MeOH)



Peak #	Ret time [min]	type	Width [min]	Area [mAU*s]	Hieght [mAU]	Area (%)
1	2.909	BV	0.0715	91.31248	19.17892	0.3981
2	3.187	VB	0.2078	12.09558	7.67903e-1	0.0527
3	6.966	VB	0.1851	12.39197	1.01012	0.0540
4	7.824	BB	0.2221	53.52818	3.54691	0.2334
5	9.909	BV	0.4541	24.83895	7.06635e-1	0.1083
6	10.380	VB	0.2468	26.39601	1.63143	0.1151
7	11.069	BB	0.2399	21.05671	1.36519	0.0918
8	12.080	BB	0.2876	6.79593	3.20172e-1	0.0296
9	12.749	BB	0.2136	5.41835	3.86408e-1	0.0236
10	13.949	BB	0.2962	19.22027	9.73984e-1	0.0838
11	14.460	BB	0.2564	5.57999	2.87556e-1	0.0243
12	15.748	BB	0.3258	2.25136e4	1068.66663	98.1548
13	19.988	BB	0.4987	144.60324	4.33994	0.6304

Totals : 2.29369e4 1103.18180

