



### Chromatogram and Results

#### Injection Details

Catalog No:	CS-W005241 Batch#Lc0419080	Run Time (min): 18.00
Vial Number:	GB6	Injection Volume: 5.00
Injection Type:	Unknown	Channel: UV_VIS_1
Comment:	1062789 A-RP-244	Wavelength: 214
Instrument Method:	10-90A,214NM,AB,6-1	Bandwidth: n.a.
Processing Method:	Quantitative	Dilution Factor: 1.0000
Injection Date/Time:	23/Apr/19 00:51	Acq Operator: wxt

#### Chromatogram



#### Integration Results

No.	Peak Name	Retention Time min	Area mAU*min	Height mAU	Relative Area %	Plates (USP)	S/N	Resolution (EP)
1		2.752	29.181	229.326	95.87	2930	18.1	n.a.
2		3.873	0.317	0.000	1.04	n.a.	0.0	n.a.
3		4.657	0.761	10.737	2.50	27575	0.8	4.24
4		5.132	0.179	2.439	0.59	38843	0.2	n.a.
Total:			30.438	242.501	100.00	69348.00		